


UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS


NEW BRUNSWICK NOVA SCOTIA PRINCE EDWARD ISLAND NEWFOUNDLAND AND LABRADOR
 NUNAVUT YUKON NORTHWEST TERRITORIES

MANUFACTURERS NAME: <u>Michell Instruments Limited</u>	
MANUFACTURERS ADDRESS: <u>48 Lancaster Way Business Park, Ely, Cambridgeshire, CB6 3NW, United Kingdom</u>	
PLANT LOCATIONS: <u>48 Lancaster Way Business Park, Ely, Cambridgeshire, CB6 3NW, United Kingdom</u>	
<p align="center">CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY</p> <p><input type="radio"/> A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers</p> <p><input type="radio"/> B Flanges: all flanges</p> <p><input type="radio"/> C Valves: all line valves</p> <p><input type="radio"/> D Expansion joints, flexible connections, and hose assemblies: all types</p> <p><input type="radio"/> E Strainers, filters, separators, and steam traps</p> <p><input checked="" type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters</p> <p><input type="radio"/> G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs</p> <p><input type="radio"/> H Pressure retaining components that do not fall into one of the above categories</p> <p><input type="radio"/> N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)</p>	<p align="center">TITLE OF THE STANDARD OF CONSTRUCTION</p> <p>Proprietary Michell Standards</p>
<p align="center">SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT</p> <p align="center">  </p>	<p align="center">TYPE OF CONSTRUCTION</p> <p>FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/> CAST <input type="checkbox"/> OTHER <input type="checkbox"/> DESCRIBE OTHER: Machined with glass to metal seal</p>
<p align="center">LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED:</p> <p>Easidew I.S., Transmet I.S., Easidew PRO I.S., Easidew PRO XP, SF72, SF82 & Sensor Block</p> <p>EX90435 - GLASS TO METAL SEAL DATA FOR C.R.N. ASSESSMENT, Pressure Test Report 10-03-C0158</p>	

DECLARATION:
 I MR. IAN ARNOLD (see note 3) employed by Michell Instruments Limited and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by BSI Management Systems as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: [Signature]
 Declared before me at Kings Lynn, Norfolk, UK
 This 23 day of June AD 2020
 Commissioner of Oaths or Notary Public: (sign) [Signature] **Sarah Louise Scott**
 Notary Public

USE THIS SPACE FOR OFFICIAL STAMP



This space for Regulatory Authority use
 This registration must be revalidated after ten (10) years from the date of issuance

CRN: 0F10920.2 Rev
 FID#: 14656
 Revision is CRN renewal. Glass to metal sensors 2000psig @ -20F to 158F

Notes:
 1. All fittings shall be registered in the name of the Manufacturer.
 2. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation.
 3. The declaration shall be made by the person having full authority and responsibility for the quality of the end product.
 4. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.

CRN expires July 10, 2030 - DG

BOILER PRESSURE VESSEL
 NOVA SCOTIA
 Date July 24/20
 C.R.N. 0F10920.28REV
 Dwg. as described
 Signed Joyce Perry
 1 of 1
 Part

ACT & REGULATIONS 1.0 - Fittings Rev. 1 06/2003