



QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

File LR1507

CERTIFICATE OF COMPLIANCE
(ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Michell Instruments Ltd
Address	Unit 48, Lancaster Way Business Park Ely, Cambridgeshire CB6 3NW, UK
Project Number	LR1507-4
Product	Process Moisture Analyzer
Model Number	QMA601
Ratings/Markings	Class I, Division 1, Groups B, C and D, T6 Ta = -25°C to +55°C Class I, Zone 1 AEx db IIB+H2 T6 Gb Ex db IIB+H2 T6 Gb Ta = -20°C to +55°C 24 V d.c., 5.2 A, 140 W Maximum working pressure = 200 kPa/2 Bar
Applicable Standards	CSA C22.2 No. 30 -1986 CSA C22.2 No. 60079-0:19 CSA C22.2 No. 60079-1:16 CSA C22.2 No. 61010-1:2012 ANSI/UL 60079-0 7th ed. ANSI/UL 60079-1 7th ed. FM 3600-2018 FM 3615-2018 ANSI/UL 61010-1 3rd ed.
Factory/Manufacturing Location	Same as Applicant

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Dave Adams, P.Eng.
 Manager, Hazardous Locations [Ex Equipment] Department

Signature:

Date: May 5, 2020