

FM Approvals
1151 Boston Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781-762-9375 www.fmapprovals.com



CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

MDM300 I.S. Portable Dewpoint Meter
IS / I / 1 / ABCD / T4 Ta = 50°C; IP66 – Ex90467.

Equipment Ratings:

Intrinsically safe for Class I, Division 1 Groups A, B, C and D.

FM Approved for:

Michell Instruments Limited
Ely, Cambridgeshire CB6 3NW, United Kingdom



This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

Class 3600	1998
Class 3610	2010
Class 3810	2005
ANSI/IEC 60529	2004

Original Project ID: 3036712

Approval Granted: July 15, 2010

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Approvals LLC



J. E. Marquedant
Group Manager, Electrical

15 July 2010
Date